



**Ralston Instruments**

Pressure Calibration Specialists

F-741-027

11 November, 2025



2903

## EU Declaration of Conformity

In accordance with ATEX Directive 2014/34/EU

Ralston Instruments, LLC  
15035 Cross Creek Park Parkway  
Newbury OH 44065 USA

This declaration of conformity is issued under the sole responsibility of the manufacturer.

**Product Name:** FieldLab Pressure Calibration Device  
**Model Number:** FLP1 Series  
**Product Options:** Wireless module  
Multiple pressure connections

The object of the declaration described above is in conformity with the relevant Union harmonization legislation:

ATEX Directive 2014/34/EU

Relevant harmonized standards and technical specification in relation to which conformity is declared:

EN 60079-0:2018  
EN 60079-11:2012  
EN 60079-26:2015

**EU Type Examination Certificate:** ITS14ATEX27973  
**Ex Ratings:** II 1 G Ex ia IIC T4 Ga  
Tamb: -10°C to +50°C  
IP67

**Issue Date:** 31 October 2019  
**Issuing Authority:** Intertek Italia, S.p.A.  
**Notified Body Number:** 2575  
**Issuing Authority Address:** Via Miglioli, 2/A – 20063 Cernusco sul Naviglio  
Milano, Italy

Signed for and on behalf of Ralston Instruments, LLC

Griffin D. Ralston  
CEO, Ralston Instruments, LLC  
11 November, 2025