

CE EC Declaration of Conformity

According to ISO/IEC 17050-1:2004

Manufacturer: Ralston Instruments, LLC
15035 Cross Creek Pkwy
Newbury OH, 44065 USA

We hereby declare that the following products


FieldLab Pressure Calibrator – FLP1	Quick-test Hose and Adapters – QTHA, QTFT
Hydraulic Hand Pump- HPGV, HPOV, XHGV, XHOV, QTHP-5K, QTHP-10K	Quick-test XT Hose and Adapters – XTHA, XTFT
Pneumatic Hand Pump – APOV, APGV, DPOV, DCAP	Benchtop Manifolds – MF2S, MF4S
Pressure and Vacuum Hand Pumps – DV0V, DPPV, DCAP-PV	Pressure Regulators – SREG, XREG
Calibration Manifolds – QTCM, QSCM	Fluid and Moisture Traps – NPAK-TRSO
Volume Controllers - QTVC	Sampling Pumps – SPOV
Nitropak Nitrogen Sources – NPAK	Quick-test Valves and Verniers
F-35 Strut Fill Assemblies – F-35	LC10, LC20 and LC30 pressure, temperature gauges

Comply with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly.

EN 97/23/EC - Pressure Equipment Directive (PED)

Product and Pressure Range	Description
Maximum Allowable Pressure (PS) ≤ 200 bar (2900 psig)	All above products for use on gases or liquids less than or equal to 200 bar (2900 psig) are classified as Sound Engineering Practice (SEP)
Maximum Allowable Pressure (PS) > 200 bar (2900 psig)	All above products for use on gases or liquids greater than 200 bar (2900 psig) on Class 1 and Class 2 gases or liquids, Ralston Instruments maintains a technical file in accordance with Annex III, Module A (internal production control)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. Ralston Instruments complies with all essential requirements of the Directives.



Griffin D. Ralston Director of Engineering
Ralston Instruments

1 August, 2018