

# CE EC Declaration of Conformity

According to ISO/IEC 17050-1:2004

**Manufacturer:** Ralston Instruments, LLC  
15035 Cross Creek Pkwy  
Newbury OH, 44065 USA

We hereby declare that the following products


Pneumatic Hand Pumps – DV0V, DPOV, DPPV, DCAP, DCAP-PV	Custom Microbore hoses with ends 1M, 2M, 1P, 1S, 2G, 2L, 2M, 2S, 2T, AA, QS, QT, XT, 2J, L4
Hydraulic Hand Pumps - QTHP, XTHP	Benchtop Manifolds – MF2S, MF4S
Calibration Manifolds – QTCM, QSCM	Pressure Regulators – SREG, XREG
Volume Controllers – QTVC, QTVC-EN	Fluid and Moisture Traps – NPAK-TRS0
Nitrogen Sources – NPAK	Sampling Pumps – SPOV
Quick-test Hose and adapters– QTHA, QTFT	Valves and Verniers – QTHA, QTVF, F35
Quick-test XT Hose and adapters – XTHA, XTFT	FieldLab Pressure Calibrator – FLP1
USN Hose and adapters – USN0	Pressure, temperature gauges - LC10, LC20, LC30

Comply with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly.

**Pressure Equipment Directive (2014/68/EU)**

Product and Pressure Range	Description
Maximum Allowable Pressure (PS) ≤ 200 bar (2900 psig)	All above products for use on gases or liquids less than or equal to 200 bar (2900 psig) are classified as Sound Engineering Practice (SEP)
Maximum Allowable Pressure (PS) > 200 bar (2900 psig)	All above products for use on gases or liquids greater than 200 bar (2900 psig) on Class 1 and Class 2 gases or liquids, Ralston Instruments maintains a technical file in accordance with Annex III, Module A (internal production control)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. Ralston Instruments complies with all essential requirements of the Directives.



Griffin D. Ralston  
Director of Engineering

22 September, 2020