

Declaration of Conformity

According to ISO/IEC 17050-1:2004

Manufacturer's Name:	Ralston Instruments, LLC	
Manufacturer's Address:	15035 Cross Creek Parkway	
	Newbury OH 44065 USA	

Declares under sole responsibility that the product as originally delivered

Product Name:	FieldLab Pressure Calibration Device
Model Number:	FLP1 Series
Product Options:	Wireless module, multiple pressure connections

Complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

EMC Directive 2004/108/EC EN 61326-1 Edition 2.0 2012-07

And conforms with the following product standards

Standard	Description	Performance Criterion	Status
CISPR 11:2009, +1:2010 Group 1	Radiated Emissions	B @ 10m	PASS
CISPR 11:2009, +A1:2010 Group 1	AC Mains Conducted Emissions	N/A	N/A
IEC61000-4-2:2008	Electro-Static Discharge Immunity test	В	PASS
IEC 61000-4-3:2006 +A1:2007, +A2:2010	Radiated, Radio Frequency, Electromagnetic Immunity	А	PASS
IEC 61000-4-8:2009	Power Frequency Magnetic Field Immunity Test	А	PASS

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. Ralston Instruments complies with all essential requirements of the Directives.

DRE

Griffin D. Ralston Authorized Representative Ralston Instruments

15 August, 2015