



Declaration of Conformity

In Accordance with ATEX Directive 94/9/EC

Manufacturer's Name: Ralston Instruments, LLC
Manufacturer's Address: 15035 Cross Creek Parkway
Newbury OH 44065 USA


Declares under sole responsibility that the product as originally delivered

Product Name: FieldLab Pressure Calibration Device
Model Number: FLP1 Series
Product Options: Wireless module
Multiple pressure connections

Complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly.

ATEX Directive 94/9/EC

And conforms with the following product standards:

 II 1 G Ex ia IIC T4 Ga
-10°C ≤ Ta ≤ 50°C IP67

Standard	EHSR
EN 60079-0:2012	Harmonized
EN 60079-11:2012	Harmonized
EN 60079-26:2007	Harmonized

EC Type Examination Certificate: Intertek Testing Services NA, Inc.
Report Number: ITS14ATEX27973
Issue Date: 2015-08-05

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. Ralston Instruments complies with all essential requirements of the Directives.

Griffin D. Ralston
Authorized Representative
Ralston Instruments

15 August, 2015