


Ralston DV0V Vacuum Test Pump Operation Manual



For all models of the Ralston DV0V Vacuum Test Pump










Table of Contents

Specifications	2
Requirements.....	3
Important Safety Notices	4
California Proposition 65 Warning.....	4
DV0V Vacuum Test Pump Overview	5
Setting Up.....	6
Calibration.....	10
Venting System.....	13
Storage and Transport	14
Maintenance.....	15
Troubleshooting	16
Support.....	17

Specifications

Vacuum Range	0 to 23 inHg (0 to 584 mmHG)
Media	Air
Vacuum Reference Port	Male Ralston Quick-test™ inlet port, no Check Valve, brass
Vacuum Inlet Port	Male Ralston Quick-test™ inlet port with cap and chain, brass
Temperature Range	0 to 130 °F (-18 to 54 °C)
Seal Materials	Buna-N, Delrin, Teflon
Construction	Anodized aluminum, brass, stainless steel
Fine Adjust Resolution	+/- 0.01 psi (+/- 0.7 mbar)
Weight	1.79 lb (0.8 kg)
Dimensions	H: 3.63 in (9.22 cm) W: 8.5 in (21.59 cm) D: 3.63 in (9.22 cm)

Requirements



* ralstoninst.com/adapters

What you need to use your DV0V Vacuum Test Pump:

1. Device Under Test
2. Wrench
3. Thread Tape
4. Ralston Quick-test™ Adapters
5. Ralston Quick-test™ Hose
6. Pressure Reference

Important Safety Notices

⚠ WARNING: Do not exceed Maximum Working Pressure for this product or damage may result.

⚠ WARNING: Device under test should be isolated from the process, vented and vent valve closed prior to use.

⚠ WARNING: Do not attempt to operate this pump until you have read and fully understand the instructions and hazards of the product.


- Any modifications to this product with custom parts can result in hazardous operation of the hand pump.
- Use eye protection while using this product. Leaking gas, parts or hoses can be ejected at high speed and may cause injury.

California Proposition 65 Warning


⚠ WARNING: Products containing Brass can expose you to chemicals, including lead, which is known to the State of California to cause cancer and birth defects or other reproductive harm. For more information, visit www.P65Warnings.ca.gov.

DV0V Vacuum Test Pump Overview

Back



Front




1. Pumping Piston
2. Vacuum Reference Port
3. Bleed Valve
4. Vacuum Inlet Port
5. Fine Adjust Piston
6. External Check Valve
7. Internal Check Valve

Setting Up


Connecting Reference Gauge

Male NPT Reference Gauge




1. Reference Gauge with NPT male connection
2. NPT Female Ralston Quick-test™ Gauge Adapter
3. Ralston Quick-test™ Hose
4. NPT Female Ralston Quick-test™ Adapter

or




Male BSPP Reference Gauge




- 1. Reference Gauge with BSPP male connection
- 2. BSPP Washer
- 3. BSPP Female Ralston Quick-test™ Adapter
- 4. Ralston Quick-test™ Hose
- 5. BSPP Female (RG) Ralston Quick-test™ Adapter

or




Female NPT Pressure Reference Gauge




1. Reference Gauge with NPT female port
2. NPT Male Ralston Quick-test™ Gauge Adapter
3. Ralston Quick-test™ Hose
4. NPT Male Ralston Quick-test™ Adapter

or



Connecting Device Under Test (DUT)




1 Isolate the Device Under Test (DUT) from the process and vent DUT prior to connecting to it.





1. Device under test (DUT)
2. NPT Male Ralston Quick-test™ Adapter
3. Ralston Quick-test™ Hose

Calibration

Prepare the Pump




Increase Vacuum




 Test point

1





2



Repeat steps 1 through 2 for each test point up-scale.

Decrease Vacuum



1

A line drawing showing a hand operating the bleed valve on the vacuum gauge. The gauge is mounted on a manifold. A curved arrow indicates the hand turning the bleed valve. A test point icon points to the bleed valve.

Slowly bleed to just above the test point. Then, close Bleed Valve.


2

A line drawing showing a hand fine-adjusting the vacuum gauge. The gauge is mounted on a manifold. A curved arrow indicates the hand turning the fine-adjustment knob. A test point icon points to the gauge's scale.

Fine-adjust to exact test point.

 Repeat steps 1 and 2 for each test point down-scale.


Venting System




When finished testing, open the Bleed Valve and vent the system.

Storage and Transport


1



2



3



* Ralston DP0V, DV0V or DPPV Test Pump Carrying Case (DP0V-CASE). Sold separately.

Maintenance

Maintenance Interval

Every 300 uses or 3 months

Maintenance Procedure

- Lubricate internal Check Valve by removing pumping piston and internal Check Valve. Squirt 2 ml of oil into the threaded end of the Check Valve and reassemble. When the pumping piston is pumped the oil will be pumped through the internal Check Valve.
- Lubricate external Check Valve by removing external Check Valve. Squirt 2 ml of oil into the threaded end of the Check Valve and reassemble. When the pumping piston is pumped the oil will be pumped through the external Check Valve.
- Lubricate the Ralston Quick-test™ fittings by squirting 2 ml of oil inside the connection.
- Lubricate the pumping piston and the fine adjust piston O-rings with silicone lubricant.

Troubleshooting

The pump will pull a vacuum, but vacuum slowly decreases

If the pump will pull a vacuum, but vacuum slowly decreases, then there is an external leak. Follow these instructions to locate and repair the leak:

1. Connect the pump to a Device Under Test (DUT) with a Ralston Quick-test™ hose.
2. Make sure the process connections are assembled wrench-tight.
3. Pull a vacuum with the Hand Pump.
4. Spray soapy water or leak detection fluid where leaks are suspected.
5. Observe where bubbles are getting sucked into the pump to determine where there is a leak.
6. Remove the leaking part and remove the O-ring.
7. Clean and lubricate the O-ring.
8. Replace the O-ring and reassemble.

Vacuum decreases when pumping piston returns to pump

If the vacuum decreases when pumping piston returns to pump, then the Internal Check Valve is not functioning properly. Follow the instructions to replace the internal Check Valve:

1. Remove the Pumping Piston.
2. Remove the Internal Check Valve.
3. Clean and lubricate the Check Valve.
4. Reinstall the Check Valve.
5. Reinstall the Pumping Piston.

When the Pumping Piston is pumped, vacuum does not increase and air is felt getting sucked into the External Check Valve located opposite of the Bleed Valve.

If when the Pumping Piston is pumped, vacuum does not increase and air is felt leaking into the External Check Valve located opposite of the Bleed Valve, then the External Check Valve is not functioning properly. Follow instructions to replace the External Check Valve:

1. Remove the External Check Valve
2. Clean and lubricate the External Check Valve.
3. Reinstall the External Check Valve.

The pumping piston is difficult to pump.

If the pumping piston is difficult to pump after years of use, then apply a thin coat of graphite grease, such as Dow Corning® Moly-kote G-n Metal Assembly Paste (or equivalent).

If the issue was not resolved by these troubleshooting instructions, then please contact support listed on page 17.

Support

Hours: **8:30 am – 5:00 pm EST**

Phone: **1 440-564-1430 • Toll Free: 1 800-347-6575 (US and Canada)**

Web: **ralstoninst.com/support**

Email: **support@ralstoninst.com**

Parts and Service: **ralstoninst.com/dv0v**

Ralston DV0V Vacuum Test Pump Operation Manual

For all models of the Ralston DV0V Vacuum Test Pump



ralstoninst.com

Hours: 8:30 am – 5:00 pm EST

Phone: 1 440-564-1430

Toll Free: 1 800-347-6575 (US and Canada)

Support: ralstoninst.com/support • Parts and Service: ralstoninst.com/dv0v

Email: support@ralstoninst.com